



नेपाल सरकार
स्वास्थ्य तथा जनसंख्या मन्त्रालय
स्वास्थ्य सेवा विभाग

फोन नं. ४५५२४२१
फ्याक्स नं. ४५५२३७५
इमेल: nphl@nphl.gov.np

राष्ट्रिय जनस्वास्थ्य प्रयोगशाला



पत्र संख्या २०८०/०८९
चलानी नं.

मिति: २०८०।११।२८

विषय: शिलबन्दी दरभाउ पत्र पेश गर्ने सम्बन्धमा ।

श्री

..... ।


प्रस्तुत विषयमा यस प्रयोगशालामा प्रयोग हुने तपसिलमा उल्लेखित किट्स, केमिकल्स खरिद गर्नुपर्ने भएकोले उक्त किट्स, केमिकल्स के कति दररेटमा उपलब्ध गराउन सकिने हो सोको दर रेट खुलाई शिलबन्दी लाहछाप गरी Eligible Criteria (अध्यावधिक संस्था/व्यवसाय दर्ता प्रमाणपत्र, भ्याट/प्यान दर्ता प्रमाणपत्र, स्वघोषणा, Power of Attorney आ.व. ०७९।८० को करचुक्ता प्रमाणपत्रको प्रतिलिपी संलग्न राखी) मिति २०८०।१२।०७ गते १२:०० बजेसम्म आइपुग्ने गरी पेश गर्नुहुन अनुरोध छ । साथै उल्लेखित किट्स, केमिकल्सको स्पेसिफिकेसन समेत यसैसाथ संलग्न राखी पठाईएको छ । उक्त शिलबन्दी दरभाउ पत्र मिति २०८०।१२।०७ गते दिउँसो २:०० बजे खोल्ने भएकोले प्रतिनिधि समेत तोकिएको पठाईदिनु हुन अनुरोध छ ।

मूल्यांकनका शर्तहरू:

१. मूल्यांकन गर्दा Eligible Criteria and Technical Specification पुरा भएका र किट्स केमिकल्सको Specification मा उल्लिखित विवरण खुल्ने कागजात पेश भएका फर्महरूलाई सारभूत रूपमा प्रभावग्राही मानिनेछ ।
२. सारभूत प्रभावग्राही (Eligible Criteria and Technical Specification) फर्महरूको मात्र आर्थिक प्रस्ताव मूल्यांकन गरिनेछ ।
३. सि.नं.१ देखि ७ सम्मका किट्स केमिकल्सको छुट्टै छुट्टै दररेट उल्लेख गर्नुपर्नेछ । सो मध्ये सबैभन्दा घटी प्रति इकाई दररेट भएको फर्महरूसँग सम्झौताको लागि पत्राचार गरिनेछ ।
४. सारभूत रूपमा प्रभावग्राही फर्महरूसँग प्रचलित खरिद कानुन बमोजिम कार्यसम्पादन जमानत लिइ सम्झौता गरिनेछ ।

तपसिल:

S.NO	Name	Unit	Qty	Rate	Total	Remarks
1	EGFR mutation PCR kit	Test	48			
2	Blood DNA extraction kit	Test	240			
3	Blood RNA extraction kit	Test	96			
4	FFPE Genomic DNA extraction kit Manual	Test	48			
5	HR- HPV DNA extended genotyping kit	Test	384			
6	BCR/ABL1 translocation qualitative PCR Kit	Test	48			
7	HLA B27 PCR Kit	Test	192			


(रमेश राज भट्ट)
निर्देशक



आरंभिक
संरक्षण



आरंभिक दुरी
कायदा गरी



हेलाथेस्टा
सम्बन्धमाथी हात धोउ

जनस्वास्थ्यका मापदण्डहरूको कडा रूपमा पालना गरौं, गराऔं र
कोभिडको संक्रमणबाट बचाऔं बचाऔं

Government of Nepal
Ministry of Health and Population
Department of Health Services
National Public Health Laboratory
Teku, Kathmandu

Program: Procurement of Kits, Chemicals, Accessories

Tender Notice No. :

Tender No. :

Fiscal Year: 2080/081

Package No.

S.N.	Purchaser's Specifications		Bidder's Compliance Sheet		
			No	Page No. in Catalogue	Remarks
	Name	Specification	UNIT		
1	EGFR mutation PCR kit	EGFR mutation detection by Real-time PCR, Total mutations covered: 28 at least, Exons covered : 18,19,20,21. Mutations included: L858R, T790M, 3 insertions Exon 20 insertions, 19 Exon 19 deletions, L861Q, S768I, G719X, fluorochrome Compactibility: BioRad CFX 96, Internal control: Endogenous human gene, 1% mutant allele in the background of 99% wild type or better sensitivity, CE IVD or USFDA, 12 months or longer Self life. 6 months of minimum expiry date.	Test		
2	Blood DNA extraction kit	Technology: Silica membrane, Format: Spin column, Capacity: Mini Spin column, Certification: CE IVD marked or USFDA approved, Processing: Manual, centrifugation, Application: Real-time PCR, Elute Volume: ~50-100 ul, Sample Type: EDTA whole blood, Output: Pure genomic DNA	Test		
3	Blood RNA extraction kit	Quality certificate of analysis of supplied lot could be traced online or submitted along with the kit, Technology: Silica membrane, Format/capacity: Mini spin column, CE IVD marked or USFDA approved, Processing: Manual Centrifugation, Application: Real-time PCR, Elution Volume: ~50-100 ul, Sample Volume: EDTA whole blood	Test		



Scanned with CamScanner

[Handwritten Signature]



Scanned with CamScanner

4	FFPE Genomic DNA extraction kit Manual	FFPE Genomic DNA extraction kit, Manual, Quality certificate of analysis of supplied lot could be traced online or submitted along with the kit, Technology: Silica membrane, CE IVD marked or USFDA approved, Format/capacity: Mini spin column, Processing :Manual, centrifugation, Application: Real-time PCR, Elution Volume: ~50-100 ul, Sample Volume: Formalin fixed paraffin embedded tissue samples.	Test		
5	HR- HPV DNA extended genotyping kit	Technology: Real-time PCR, Genotypes covered: at least 14 common High risk Genotypes, Genotyped separately: HPV 16, 18, 31, 33, 35, 39, 45, 51, 52, 53, 56, 59, 66, 68 , Format: Multi Tube and Multiplex, Compactibility: BioRad CFX 96,, Endogenous Human Gene as Internal Control, 2000 copies/ml or better sensitivity, CE IVD or USFDA certification, Stability of 12 Months, minimum of 6 months of expiry.	Test		
6	BCR/ABL1 translocation qualitative PCR Kit	Real time PCR format, should detect and differentiate Major, Minor and Micro breakpoints. Positive and negative controls shall be included. CE IVD certified	Test		
7	HLA B27 PCR Kit	Human Leukocyte Antigen (HLA) B27 Detection kit by Real-Time (RT)-PCR, Compactibility: Riorad CFX-96, ABI Quantstudio 5, ABI 7500 Fast, Endogenous Human Gene as Internal Control, Stability of 12 Months, minimum of 6 months of expiry. CE IVD Certification.	Test		

